



## CLEANING GUIDE



# LQS FLOOR TRACK SYSTEM



Read and understand all of the instructions and information in this document before cleaning this product.

**SAVE THESE INSTRUCTIONS**

## TO CLEAN THE ALUMINUM COMPONENTS

- Dampen a clean cloth with Simple Green cleaner (or any citric or acetic acid cleaner) on the aluminum components to wipe them clean. DO NOT spray the cleaner directly onto the aluminum, but spray/wet the cloth rag and wipe the aluminum components with that solution
- If you are looking to restore the shine on the components, simply mix equal parts water and white vinegar and wipe the aluminum surface with a cloth rag soaked in that solution

## TO CLEAN THE ALUMINUM TRACK

- With a clean, DRY cloth, gently wipe the inside of the track to remove any dust and debris
  - After wiping the track out with a DRY cloth, you can vacuum any remaining debris out of the track (if in an SPD area the vacuum must have a HEPA filter)
- 

## LIMITED 5-YEAR WARRANTY

LogiQuip products are covered by a Limited 5-Year Warranty from date of receipt against defect in materials or workmanship under normal use and service, excluding Electronic Components, Cart Covers, Custom Products, and LogiCell Tambour Doors. During this period, LogiQuip will, in its sole discretion, repair defects in materials or workmanship that existed when the product was received or replace the product. Visit [LogiQuip.com/Warranty](https://www.logiquip.com/Warranty) for more details.

## ABOUT US

From quality products and services to patented, industry-leading innovations, LogiQuip has been providing smart, future-focused solutions exclusively to the healthcare industry since 1992. Our knowledgeable representatives partner with you to understand your needs, then design and implement the best solution for any space in your facility. With a focus on infection control, lean processes and ROI justification, our extensive line of products and services will help make your facility more effective and efficient now and for years to come.